# Single-Phase Power Meter 7110/7120

## Features

- AC / DC Dual amp/watt-hour meter
- Wide range 0.001W-16KW
- Connection software attached
- Standby Power D.P.I. of 0.001W
- With crest factor ratio display
- Up to the 50 levels harmonic wave analysis capability
- 1000 sets of measurement data storage space
- Current crest factor is highest CF9





CE RS-232⊠ GPIB⊡

## **Accessories / Fixtures**

## Standard

- Power Cord
- RS232cable
- F71201 TEST BOX
- TL218 Alligator Clips
- TL208 2mm Test Probe





- GPIB cable



## Specification

Model Name	7110	7120
Frequency Measurement Mode	To achieved stable base frequency measurement (variation less than 1%) by voltageor current (non- inverter)	
Frequency Range	DC15Hz - 10kHz DC15Hz - 100kHz	
Data Length	Dual 4096×16 RAM for voltage & current	
ADC Resolution	16 bits	
Sampling Rate	AC 50Hz/60Hz basic sampling rate 100 KSPS /120 KSPS	
Arithmetic Precision	Watt/VRMS/IRMS/MEAN/PF/Deg/Line filter 32bits	
Frequency Filter	500Hz cut off, digital chip	filter based on 25MHz
Signal Filter	500Hz-3db digital filter based on Butterworth 50Hz-0.03% reading, 60Hz-0.05% reading	
Frequency Acquisition Mode	Voltage / current 100MHz meter chip	baseband digitaldynamic
Phase Lead Detection	Subject to the current, analog/digital hybrid detecting (error less than 5 degrees)	

## Range

Current (fixed/auto)	0.01A, 0.03A, 0.1A, 0.3A, 1A, 3A, 10A, 20A
Voltage (fixed/auto)	10V, 30V, 100V, 300V, 600V

# Specification

Model Name		7110	7120		
Power Supply		Voltage 100 ~ 240Vac Frequency 50/60Hz			
Display		Seven-segment display			
Interface		RS-232	RS-232+GPIB		
Flash Memory		6 Sets			
Environment		Temperature: 23°C±5°C, Humidity: 20-80%RH			
Dimension (W*H*D)		227×101×300 mm			
Weight	Weight		1.85kg		
Measurement bandwidth	ı	DC 15Hz-10kHz	DC 15Hz-100kHz		
Harmonic (option)		Yes/NA	Yes/NA		
Model		7110-10k-HARM 7110-10k	7120-100k-HARM 7120-100k		
Fixture		F71201 TEST BOX			
Harmonics					
Analysis base	To achieved stable fundamental frequency analysis by voltage or current (non-inverter)				
Frequency Range	45F	Iz - 440Hz			
FFT Data Length	1024				
FFT Data Format	32 bits				
MeasurementProjects	<ul> <li>1-50 THD, 1-50 level voltage and current V [n], A[n]</li> <li>1-50 level voltage and current distortionpercentage</li> <li>V [n%], A [n%]</li> <li>1-50 level watts W [n]</li> <li>1-50 level watts distortion percentage Watt W [n%]1-</li> <li>50 level voltage and current angle DEG [n] Vrms, Irms, Watt, PF</li> </ul>				

## **Parameters Measurement Range**

Vrms	0.1V-600V	PF	0.000-±1.000
Vdc	0.1V-600V	Deg	-180°-+180°
Irms	0.1mA-20A	THD	0.01%-999.99%
Idc	0.1mA-20A	Hz	15Hz-100kHz
W	0.01W-16kW		

## RMS/MEAN Mode Voltage & Current Accuracy (23°C $\pm 5^{\circ}C$ )

$15Hz \le f < 45Hz$	$\pm (0.1\% \text{ of reading} + 0.4\% \text{ of range})$
$45Hz \!\leq\! f \!\leq\! 66Hz$	$\pm (0.1\% \text{ of reading} + 0.1\% \text{ of range})$
$66\text{Hz} < f \le 1\text{kHz}$	$\pm (0.1\% \text{ of reading} + 0.2\% \text{ of range})$
$1 \text{kHz} < f \le 10 \text{kHz}$	$\pm (0.07*f\% \text{ of reading} + 0.3\% \text{ of range})$
$10 \text{kHz} < f \leq 100 \text{kHz}$	$\pm(0.5\% \text{ of reading} + 0.5\% \text{ of range}) \pm[\{0.04 \times (f-10)\}\% \text{ of reading}]$

F unit is 1KHz

When the L-FILTER sets as ON: 45Hz~66Hz frequency range allowable error-0.03 %~-0.05 of reading

When the AC is measured, if the fundamental frequency exceeds 200Hz, the F-Filter is required to be turned off in order to measure the mostaccurate value

\*\*When the frequency range is more than 10KHz, the 7120 starts to support

## DC Mode Voltage & Current Accuracy (23°C ±5°C )

10V-600V	$\pm 0.2\%$ reading $\pm 0.2\%$ of range	0.01A-20A	$\pm$ {(0.2)% of reading + 0.2 % of range} $\pm$ offset
To add up the OFFSET errors of various files during measuring the DC current			

## Power (W) Accuracy (23°C ±5°C )

Maximum Power (W) value is determined by the highest range of voltage profile		
DC $\pm 0.2\%$ reading $\pm 0.5\%$ of range		
$15Hz \le f < 45Hz \pm (0.3\% \text{ of reading} + 0.2\% \text{ of range})$		
$45$ Hz $\leq f \leq 66$ Hz $\pm (0.1\% \text{ of reading} + 0.1\% \text{ of range}) 66$ Hz		
$< f \le 1 kHz \pm (0.2\% of reading + 0.2\% of range)$		
1kHz < f $\leq$ 10kHz $\pm$ (0.4% of reading + 0.3 % of range) $\pm$ [{0.06×(f)}% of reading]		
$10 \text{kHz} < f \le 100 \text{kHz} \pm (0.5 \% \text{ of reading} + 0.5 \% \text{ of range}) \pm [\{0.09 \times (f-10)\}\% \text{ of reading}]$		
Incidental Allowable Error Conditions		
Signal Filter Error (AC) Frequency between 45-66Hz: Add 0.3% of reading. Frequency between45-66Hz: Add 1% of reading beyond		
CF9 Error (DC) Add range tolerance * 3		
Accuracy Effect of the Phase Error of the Power		
When the power factor PF is 0, the error range of Watt is		
Situation 1: for 45Hz < f, Add±1.0% of VA		
Situation 2: for $45$ Hz > f or f > $66$ Hz		
Add $\pm$ {(3.5 + 0.5×f)% of VA} for up to 100kHz as reference dataThe		
unit for frequency f is kHz.		
When the power factor is $0 < PF \leq error range$		
When $0 < PF \le 1$ ( $\theta$ : phase angle of the voltage and current) for 45Hz $\le f \le 66Hz$ . Add $\pm$ power reading *{tan( $\theta$ )*(0.5)}%		
for f < 45Hz, f > 66Hz. Add $\pm$ power reading *{ tan $\theta$ *(0.5×f+0.2) }%		
Error within 12 months Add ±(0.5% of reading)		

GMGA MEASURING No. 33 Alley 99/120 Dinh Cong Ha, Dinh Cong Ward, Hoang Mai District, 10000 Hanoi City, Vietnam Telephone: <u>+84 845 969 336</u> Email: <u>info@gmga.vn</u> Website: <u>https://gmga.vn/</u>